

1. मिसिल संख्या 6-74/09- एन.आर.सी.पी.बी / 8418-10

दिनांक:
11 / 12/09

उपकरण का नाम कम्प्यूटर्स खरीदने हेतु

अंतिम तिथि 29-12-09

Specification

Branded Computer with the following specifications

Intel(R) Core(TM) i7-920 Processor (2.66GHz, 8MB), Genuine Windows Vista(R) Home Premium 64 bit SP1 Edition (English), Microsoft(R) Office Home & Student 2007 OEM Version, Dell S2409W 24 " Full HD Widescreen Flat Panel Monitor, 3GB (3X1GB) DDR3 SDRAM 1066MHz Memory, 500GB SATA 3.0Gb/s Hard Drive with Native Command Queuing, Seagate Expansion External Drive (2nos.) 2x1 TB, Blu-ray BD Combo Drive (DVD/CD burner with double layer with Blu-Ray Reader), 512MB ATI(R) Radeon(TM) HD 4850, Integrated 7.1 Audio, Dell(TM) 2.1 Multimedia Desktop Speaker System, Dell(TM) Consumer Wireless Keyboard and Mouse Bundle, Adobe Reader (English) studio XPS Base - Direct, studio XPS 435MT Resource DVD, 2x1nos. USB lights and 2x1nos. USB fans, 19-in-1 Hi-Speed Media Reader with Bluetooth(R) 2.0, Basic System Installation

One No

Microtek 800VA UPS, double Power, double battery-

Note: please submit EMD an account of Rs 1800- in the shape of Bank draft/ pay order in favour of Project Director, NRCPB, without EMD quotation will not be accepted.

One No

2. मिसिल संख्या 6-113/09- एन.आर.सी.पी.बी / 8443-15

दिनांक:
11 / 12/09

उपकरण का नाम माईक्रोस्कोप (फेस कोन्ट्रैस्ट स्टीरियोजूम के सा(०)) खरीदने हेतु

अंतिम तिथि 29-12-09

Specification

Microscope (phase contrast with stereozoom)

OPTICS	Parallel-OPTICS ZOOM SYSTEM.
MAGNIFICATION	8x-80x (with Eyepiece of 10X and Objective of 1X)
Eyepiece Tube	Trinocular eyepiece tube with inclination angle of 20 degrees interpupillary distance adjustment 48-75 mm
Eyepieces	10x (F. O.V.22)
Zoom Range	0.8x-8.0x
Zoom Ratio	10.1
Objectives	Plan Apo 1x
Working Distance	70mm
Stand	Diascopic Stand (6V-20W) with built-in power supply with anti-backlash focusing
Episcopic Illumination	Led ring Illuminator
Digital camera system	Digital camera capable of handling bright field, Fluorescence, DIG, dark field images with 2/3" high density CCD chip, 5 mega pixel resolution. Live display mode with roi & binning; binning modes: 2x2, 4x4, digital zoom: upto 16x (8 steps); interval shooting: 5 sec.-12 hr intervals. USB port for attaching camera onto desktop/laptop through single wire.

One No

	Software should come along with camera for acquiring & capturing of images, white balance, contrast, brightness & gamma correction, annotation etc.
Optional Accessories	Plan 2X Objective
Note:	Computer workstation with UPS has to be Provided for Imaging, Capturing and data Processing Microscope and camera should be from same manufacturer for best compatibility.

Note: please submit EMD an account of Rs 9000- in the shape of Bank draft/ pay order in favour of Project Director, NRCPB, without EMD quotation will not be accepted.

3 मिसिल संख्या 6-114/09- एन.आर.सी.पी.बी / 8428-15

दिनांक:
11 / 12/09

उपकरण का नाम
अंतिम तिथि
Specification
माईक्रोसेंटीफ्यूज खरीदने हेतु
29-12-09
Micro centrifuge
Technical Specifications

One No

- 1010-2-020 safety Compact centrifuge for speed up to 14,500 rpm
- Rotor for 12 X 1.5 ml & 2 ml tubes
- G value should be more than 14,000
- Timer from 15 sec to 99 minutes, with continuous function.
- Adaptors for 0.2, 0.5 ml tubes
- Acceleration to reach max speed in 13 sec
- Breaking time max. speed 12 sec.
- Brushless motor
- PTF coated organic chemicals- resistant rotor & lid with knurled nut.
- Weight less than 5 kg.
- Bore size should be 11 mm.
- Fixed angle rotor should be at 45°
- Should confirm to IEC regulations.
- Separate short spin button.
- IVD certified
- Should have option for PCR strip rotor.

Note: please submit EMD an account of Rs 2000- in the shape of Bank draft/ pay order in favour of Project Director, NRCPB, without EMD quotation will not be accepted.

4 मिसिल संख्या 6-110/09- एन.आर.सी.पी.बी / 8346-15

दिनांक:
09 / 12/09

उपकरण का नाम
अंतिम तिथि
Specification
कम्प्यूटर्स टेबल खरीदने हेतु
29-12-09
Computer Tables (Good Quality with Latest Features)

Note: Please Submit EMD an account of Rs. 1000/- In the shape of Bank Draft/ Pay Order in favour of Project Director, NRCPB without EMD Quotation will be not be accepted. 4 No